



Urbana Fire Department

400 South Vine Street
Urbana, Illinois 61801
(217) 384-2420-Phone
(217) 384-2449-Fax

The following questions have been submitted for RFP 1819-16.

1. Can you email me the specifications?

A: The specifications are that of the Phillips FR3 or equivalent.

2. Are you aware of the manufacturer backorder on the FR3

A: Yes, we are aware of the manufacturer backorder on the FR3

3. By graphical display, are you meaning an LCD screen that provides written prompts or just the images for conduction CPR?

A: The graphical display must display the ECG.

4. Where are items and pricing to be input for this RFP? I do not see an area to input product information on the solicitation.

A: Please price your devices and items and place in a sealed envelope or email daodle@urbanaininois.us with RFP 1819-16 in the subject line.